

Calibrator for automated systems

10759350360 12 x 3 ml

LOT 391517

2021-08

Roche/Hitachi cobas c 503

Roche Calibrator code / Code d'étalon Roche: 401

Value sheet / Fiche de valeurs théoriques

Component Analyte	Method Méthode	Code	ACN	Value Valeur	Units Unités	
Lithium	Colorimetric / Colorimétrique (08057974190)	LI	20840	1.51	mmol/L	
Magnesium Magnésium	Colorimetric / Xylidyl blue, Gen. 2 (08058019190) Colorimétrique / Bleu de xylidyle Gen. 2	serum - sérum /plasma	MG2	20890	1.11	mmol/L
		urine	MG2U	20891	*	
Phosphorus Phosphore	Molybdate (UV), version 2 (08058610190)	serum - sérum /plasma	PHOS2	20990	1.71	mmol/L
		urine	PHOS2U	20991	*	
Total Protein. Protéines totales	Colorimetric Biuret (08058652190)	serum - sérum /plasma	TP2	21110	49.1	g/L
Triglycerides Triglycérides	Enzymatic Colorimetric GPO, (08058687190)		TRIGL	21130	1.59	mmol/L
Uric Acid Acide urique	Enzymatic Colorimetric Ver 2 (08058750190)	serum - sérum /plasma	UA2	21170	334	µmol/L
		Urine	UA2U	21171	*	
Urea Urée	UV / Kinetic, (08058806190)	serum - sérum /plasma	UREAL	21191	17.1	mmol/L
		Urine	URELU	21190	*	

**Transfer of calibration from serum/plasma application - Transfer de la calibration de l'application sérum/Plasma*

***Transfer of calibration from application for total cholinesterase - Transfer de la calibration de l'application pour la cholinestérase totale*

Note: A clinically insignificant difference may be seen between the value(s) listed on the value sheet and the value(s) obtained from the instrument readable data. This is caused by the rounding of value(s) during conversion from the unit in the instrument readable data to the unit that is being used. This may affect the last digit of the value. / Les valeurs listées dans la fiche de valeurs peuvent diverger de celles lues par le code-barres. Cette différence sans intérêt clinique concerne, le cas échéant, le dernier chiffre de la valeur et est due à un arrondissement des valeurs lors de la conversion de l'unité contenue dans le code-barres vers l'unité à utiliser.

All values are USA Standardized / Toutes les valeurs sont établies aux États-Unis.

1 Standardization using the reference preparation ERM-DA 470k (European Reference Material) / Standardisation à l'aide de la préparation de référence ERM-DA 470k (European Reference Material)

2 Based on reference method (modified Abell-Kendall) from Center for Disease Control, Atlanta, Georgia 30333, USA. / Selon la méthode de référence (Abell-Kendall modifiée) du Center for Disease Control, Atlanta, Georgia 30333, USA.